

Calibration

There are many types of calibration available in our industry. SCS provides what is sometimes referred to as Level 1 calibration.

What our NIST traceable calibration consists of:

- Product is calibrated to meet SCS published specifications
- A printed certificate is provided stating that we use accepted procedures and standards traceable to the National Institute of Standards and Technology
- The calibration certificate will normally consist of a model number, serial number, date, and the signature of the technician that performed the calibration
- We apply a calibration sticker to the product listing the identification number, the calibration date, and the initials of the technician performing the calibration

Our NIST traceable calibration does not include:

- It does not include any data or measurements, before or after calibration
- It does not include any NIST numbers.
- It does not include any listing of the calibration equipment used
- It is not calibration in accordance to a Mil-Standard, ANSI, or other standard
- It does not include any ambient temperature or humidity measurements
- It does not include any warranty as to length of time the calibration is valid